



Certificate of Compliance

Certificate: 2045327

Master Contract: 243657

Project: 2122872

Date Issued: 2009/01/21

Issued to: Davidge Controls

583 North Refugio Road
Santa Ynez, CA 93460
USA
Attention: Dean Davidge

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Issued by: Calvin Luong

Authorized by: Mark Christopherson, P.E.,
Manager, Certification Services

PRODUCTS

CLASS 8721 85 - ELECTRICAL EQUIPMENT FOR LABORATORY USE - Certified to
US Standards

CLASS 8721 05 - LABORATORY EQUIPMENT - Electrical

Watt-hour meter, permanently connected, rated 110-130Vac to neutral, 208-240Vac line to line, 50-60Hz, 2/3W (max).

Meters



Certificate: 2045327

Master Contract: 243657

Project: 2122872

Date Issued: 2009/01/21

Model numbers *abcd/ee* where:

a = 4 or 5

b = 0, 1 or 4 (must be 1 when *a* = 5)

c = 1 or 2

d = 0, 1, 2, 3 or 4

ee = blank, "CA", or any number less than 64

Meter Enclosures - 81111 & 81311 plus:

Model numbers *8abcnn/dd* where

a = 1, 2, 3 or 5

b = 1, 2 or 3

c = 2

nn = any number less than 25

dd = blank or any number less than 25

Note: The equipment is Class 2, Pollution Degree 2, Measurement Category II.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

UL Std. No. 61010-1 (2nd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements



Supplement to Certificate of Compliance

Certificate: 2045327

Master Contract: 243657

The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.

Product Certification History

Project	Date	Description
2122872	2009/01/21	Update report to correct component typos.
2045327	2008/07/28	Certification of wathour meter